

University of Dundee

Subcellular spatial resolution achieved for deep-brain imaging in vivo using a minimally invasive multimode fiber

Vasquez-Lopez, Sebastian A.; Turcotte, Raphaël; Koren, Vadim; Plöschner, Martin; Padamsey, Zahid; Booth, Martin J.

Published in:

Light: Science and Applications

DOI:

[10.1038/s41377-018-0111-0](https://doi.org/10.1038/s41377-018-0111-0)

Publication date:

2018

Licence:

CC BY

Document Version

Publisher's PDF, also known as Version of record

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):

Vasquez-Lopez, S. A., Turcotte, R., Koren, V., Plöschner, M., Padamsey, Z., Booth, M. J., Čížmár, T., & Emptage, N. J. (2018). Subcellular spatial resolution achieved for deep-brain imaging in vivo using a minimally invasive multimode fiber. *Light: Science and Applications*, 7(1), 1-6. [110]. <https://doi.org/10.1038/s41377-018-0111-0>

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LETTER

Open Access

Subcellular spatial resolution achieved for deep-brain imaging in vivo using a minimally invasive multimode fiber

Sebastian A. Vasquez-Lopez¹, Raphaël Turcotte^{1,2}, Vadim Koren¹, Martin Plöschner³, Zahid Padamsey¹, Martin J. Booth², Tomáš Čížmár^{3,4} and Nigel J. Emptage¹

Abstract

Achieving intravital optical imaging with diffraction-limited spatial resolution of deep-brain structures represents an important step toward the goal of understanding the mammalian central nervous system^{1–4}. Advances in wavefront-shaping methods and computational power have recently allowed for a novel approach to high-resolution imaging, utilizing deterministic light propagation through optically complex media and, of particular importance for this work, multimode optical fibers (MMFs)^{5–7}. We report a compact and highly optimized approach for minimally invasive in vivo brain imaging applications. The volume of tissue lesion was reduced by more than 100-fold, while preserving diffraction-limited imaging performance utilizing wavefront control of light propagation through a single 50- μm -core MMF. Here, we demonstrated high-resolution fluorescence imaging of subcellular neuronal structures, dendrites and synaptic specializations, in deep-brain regions of living mice, as well as monitored stimulus-driven functional Ca^{2+} responses. These results represent a major breakthrough in the compromise between high-resolution imaging and tissue damage, heralding new possibilities for deep-brain imaging in vivo.

Presently, non-invasive (surface) high-resolution imaging of brain tissue can achieve micrometer resolution up to penetration depths of $\sim 1\text{ mm}$ ¹. Beyond this limit, scattering and optical aberrations introduced by the heterogeneous refractive index distribution within brain tissue prohibit observing subcortical structures (even in mice), many of which are implicated in important neuronal processes such as memory formation and gating of sensory and motor information, as well as neurological diseases^{1–4}. The importance of visualizing these brain regions has precipitated the development of diverse optical strategies, including the removal of overlying

cortical structures⁸ and insertion of fiber bundles⁹ and graded index (GRIN) lenses^{10,11}. Unfortunately, these approaches create substantial mechanical lesions of the tissue, precipitating neuropathological responses that include inflammation and gliosis¹² and possibly ultimately compromising the physiology of neuronal networks and behavior of the animal^{13,14}. Here, we adopted an endoscopic approach utilizing multimode optical fibers (MMFs). The principles behind the MMF imaging method are detailed in the Supplementary Methods^{5–7}. The optical geometry was optimized to provide the functional stability and mobility necessary for use in vivo (Fig. 1a). Integral to the system is a liquid-crystal spatial light modulator (LC-SLM), which enabled manipulation of the propagating light field through the optical path comprising an arbitrary MMF length. Prior to the commencement of imaging, the LC-SLM was used in a calibration procedure during which we acquired a transmission matrix (TM), fully describing the light field

Correspondence: Nigel J. Emptage (nigel.emptage@pharm.ox.ac.uk)

¹Department of Pharmacology, University of Oxford, Mansfield Road, Oxford OX1 3QT, UK

²Department of Engineering Science, University of Oxford, Parks Road, Oxford OX1 3PJ, UK

Full list of author information is available at the end of the article.

These authors contributed equally: Sebastian A. Vasquez-Lopez, Raphaël Turcotte, Vadim Koren

© The Author(s) 2018



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.




Fig. 1 MMF imaging system for minimally invasive deep-brain in vivo imaging achieves diffraction-limited spatial resolution. **a** Schematic of the experimental system. (i) The source unit distributes the laser light for fluorescence excitation and calibration. (ii) The portable imaging arm controls the light propagation through the MMF by wavefront shaping using an LC-SLM. (iii) The calibration unit is used for the acquisition of the transmission matrix prior to imaging. (iv) The sample unit holds tissue sections and head-fixed mice. DM dichroic mirror, MMF multimode fiber, MO microscope objective, PMT photomultiplier tube, SMF single-mode fiber. **b** Relationship between image resolution and instrument footprint for in vivo light-based imaging modalities (Supplementary Table S1). The MMF data point is shown in blue. **c** Picture of a mouse brain (bottom) with optical components for in vivo imaging (top): a 60 \times water immersion objective (left), a GRIN lens (middle), and the 50- μm -core MMF used in this study (right)

propagation within the optical system¹⁵. The availability of the system-specific TM then allowed us to produce a set of field modulations, which were employed in the image acquisition procedure. Each of these modulations, when applied at the LC-SLM, produced a diffraction-limited spot at a specific location across the fiber output plane. Importantly, spots may be generated at an arbitrary distance from the distal fiber facet.

The source unit of the system distributes laser light (488 nm) into two single-mode optical fibers, one to deliver the excitation signal to the LC-SLM and the other to provide a reference signal during calibration. The main optical arm, comprising the LC-SLM, MMF probe, relay optics, and a fluorescence detection unit, was designed to be compact and is embedded within a robust caged framework housed on a three-dimensional micro-positioning stage to facilitate alignment during calibration as well as navigation of the MMF probe into the brain tissue

(Supplementary Methods). Finally, the calibration arm, used only during TM acquisition, relays the MMF probe output signal to a camera, where it interferes with the reference beam. The implementation of a GPU-accelerated toolbox for LC-SLM control⁷ enabled the acquisition of the full transformation matrix of a 50- μm -core MMF in <4 min. Imaging can be performed immediately after the TM acquisition, and the calibration did not change in time, provided that the fiber was not deformed and that its position relative to the rest of the optical system was not changed (Supplementary Note S1). This combination of procedures provides the basis for fiber-based volumetric point-scanning fluorescence microendoscopy, with an endoscopic probe whose diameter is several fold smaller than any used previously. Diffraction-limited performance was achieved (FWHM of the excitation PSF: $1.27 \pm 0.01 \mu\text{m}$ laterally and $20.2 \pm 1.4 \mu\text{m}$ axially; resolution of fluorescent objects from the





Fig. 2 MMF imaging of ex vivo tissue reveals dendritic spines and axonal boutons. **a** Confocal image of a hippocampal neuron in a rat brain slice. **b, c** Structural images through the MMF (**b, c** grey images) and confocal microscopy (**b, c** green images) of the same regions of the neuron in (**a**). Axonal boutons (**b**) and dendritic spines (**c**) were clearly identified using each imaging modality. Scale bars: 20 μ m

Rayleigh criterion: 1.35 μ m; both measurements at 488 nm; NA: 0.22; Supplementary Fig. S1a, b) and, importantly, was maintained up to 100 μ m from the distal facet (Supplementary Fig. S1c)—i.e., throughout the range of dynamic refocusing used in this study. This represents an enormous step in the trade-off between image resolution and device footprint (Fig. 1b, c and Supplementary Table S1).

To assess the performance of our system when imaging neuronal structures, we first conducted imaging trials in ex vivo brain slices from the rat hippocampus (Fig. 2). Fluorescently labeled neurons were imaged using both a standard confocal microscope equipped with a 60 \times water immersion objective (Fig. 2b, c green images) and our fiber-based system (Fig. 2b, c gray images). Dendritic spines and axonal boutons were clearly visible, with identical structures identifiable in both MMF and confocal images. These results suggest that the device would be suitable for structural imaging studies. We wish to stress that our digital scanning approach is free from the granular artifacts commonly observed when using computational approaches^{16,17} because these can be easily confused with synaptic structures. This is a particular concern where high-resolution images are collected from

sparsely labeled samples over a small area, as it is frequently the case in neurobiology^{18,19}. Furthermore, previous implementations of the MMF imaging system have failed at achieving sufficient spatial resolution for the visualization of dendritic spines and exclusively reported images of soma^{12,20}. Our results offer robust validation of the MMF approach for the acquisition of fluorescent images in living tissue. A limitation is that the field-of-view currently corresponds to the core diameter of the MMF (50 μ m); if necessary, a wider field-of-view can be achieved using an MMF with a larger core diameter (105 μ m) while maintaining the spatial resolution (Supplementary Fig. S2). Using MMF with a much larger diameter would increase invasiveness but may also be challenging to implement because it would require a wavefront-shaping device with many pixels because the number of input modes required scales to the square of the core diameter.

Having demonstrated that the system performs well when imaging living neuronal tissue ex vivo, we set out to explore whether we could achieve comparable results in vivo. Here we sought to image neurons from deep regions of the intact brain of live mice. We used transgenic Thy1-GFP line M mice that express fluorescently



labeled neurons sparsely throughout the nervous system, an approach commonly used for in vivo neuronal structural imaging studies (Supplementary Note S2)¹⁸. Because the diameter of our fiber was small, 125 μm inclusive of cladding, we could insert the fiber directly into the brain tissue via a small craniotomy and image in real-time as we advanced slowly through the tissue. This is a significant improvement over existing methods for deep-brain imaging with equivalent spatial resolution that require extensive surgery and aspiration of the overlying brain tissue. On identifying our target structure, we could begin imaging immediately, again in marked contrast with other endoscopic brain imaging methods that require several days post-surgery before imaging can commence¹². Critically, we saw little evidence of damage to blood vessels; thus, images were not obscured by tissue bleeding. Although chronic imaging with the MMF system was not demonstrated here, its minimal invasiveness show promise in minimizing the post-implantation recovering period.

Figure 3 shows images of a fluorescent neuron in the dorsal striatum imaged after lowering a fiber 1.8 mm into the brain of an anesthetized mouse and collecting an

image stack at different focal planes beneath the fiber (Fig. 3a, b). A dendritic branch on which there were dendritic spines was clearly identifiable (Fig. 3c) and the best in-focus axial plane identified using digital refocusing. The ability to adjust the focal plane while maintaining the fiber at a fixed position represents a further considerable advantage of the MMF fiber system for in vivo imaging. The imaging plane can be adjusted over a range of 0–100 μm from the fiber facet with no movement of the fiber and therefore no mechanical consequence for the brain tissue. This further minimizes the impact of fiber placement into the tissue. The extent to which the placement of the fiber impacts upon the neuronal tissue during image collection is shown in a post-mortem section of brain tissue (Fig. 3d). Labeled neurons and their dendritic processes remained alive and intact even when located in close proximity to the fiber tract (Fig. 3d, inset).

While structural imaging is a valuable technique, we sought to assess whether fiber-based dynamic imaging can also be achieved. Our system detected changes in fluorescence elicited in a neuron loaded with the Ca^{2+} indicator OGB-1, in ex vivo brain slices from the rat hippocampus, after exposure to potassium (45 mM;




Fig. 4a, b). The variation in fluorescence (dF/F) was calculated by subtracting the whole image taken immediately after the addition of potassium (Fig. 4b) from the one taken immediately before (Fig. 4a). For in vivo imaging, neurons of the medial geniculate body (MGB)—a part of the auditory thalamus—of C57BL/6 mice were sparsely labeled with the Ca^{2+} reporter GCaMP6m in order to record action potentials elicited in response to auditory stimuli. We exploited the digital point-scanning capability of our SLM-based system for random-access imaging of a reduced number of pixels at 33.3 Hz. Using this approach, we could reliably observe large stimulus-driven Ca^{2+} transients from sound-responsive neurons in anesthetized mice in vivo (Fig. 4b–d). In its current configuration, the imaging system can operate at a scan rate of ~ 10 ms per pixel, determined by the maximum refresh rate of 100 Hz for the LC-SLM (frame rate = 2.4 s per 120×120 -pixel frame). The LC-SLM could be combined with an acousto-optic deflector to increase the frame rate^{6,21}, but this comes at the cost of increasing the system complexity.

Digital micro-mirror devices are an alternative to LC-SLMs for spatial-light modulation^{22–24}, achieving update-rate orders of magnitude larger than LC-SLM (up to 22 kHz), and could be incorporated into our geometry.

In summary, we have achieved structural and functional in vivo fluorescent imaging of neurons within deep-brain structures of mice that we believe to be the least invasive, deep-brain, high-resolution approach reported to date. This method provides a route to achieving high-resolution optical access to deep-brain subcellular processes in living and ultimately in freely behaving animals, one of the most unique and appealing possibilities of MMF imaging systems, with minimal disruption to the associated circuitry. This study also prominently demonstrates the applications of wavefront-shaping microscopy in biomedical research²⁵ and potential future advances in minimally invasive imaging in vivo in a multitude of organs. Future developments will now aim to achieve the scanning speeds necessary for dynamic imaging over wider areas, enabling neuronal population activity to be

monitored and a critical step toward brain imaging in freely moving animals (Supplementary Note S3). Additionally, optical sectioning strategies must be devised for the imaging to be truly three-dimensional. Some progress has been made in this regard by implementing a form of holographic confocal microscopy with reflectance contrast²⁶. More suitable for fluorescence imaging would be to exploit the intrinsic optical sectioning offered by two-photon microscopy, which has been demonstrated through a step-index MMF²⁷. Finally, super-resolution imaging could be achieved by implementing a STED configuration because, unlike structured illumination or localization microscopy, the spectrally broad fluorescence would not have to be imaged through the MMF—i.e., the control of two monochromatic illumination beams would be sufficient²⁸.

Acknowledgements

M.P. and T.C. acknowledge support from the University of Dundee and Scottish Universities Physics Alliance (PaLS initiative). T.C. acknowledges support from the European Regional Development Fund, Project No. CZ.02.1.01/0.0/0.0/15 003/0000476. S.A.V.L., V.K., Z.P., R.T., M.J.B. and N.E. acknowledge support from the John Fell Fund, the BBSRC (TDRF) and the MRC (UK). We thank V. De Paola (Imperial College) for providing the transgenic Thy1-eGFP-M mouse line. We thank Véronique Vienne for technical assistance.

Author details

¹Department of Pharmacology, University of Oxford, Mansfield Road, Oxford OX1 3QT, UK. ²Department of Engineering Science, University of Oxford, Parks Road, Oxford OX1 3PJ, UK. ³School of Engineering, Physics and Mathematics, College of Art, Science & Engineering, University of Dundee, Nethergate, Dundee DD1 4HN Scotland, UK. ⁴Institute of Scientific Instruments of the CAS, Královopolská 147, 612 64 Brno, Czech Republic

Conflict of interest

The authors declare that they have no conflict of interest.

Supplementary information is available for this paper at <https://doi.org/10.1038/s41377-018-0111-0>.

Received: 26 July 2018 Revised: 3 December 2018 Accepted: 3 December 2018

Published online: 19 December 2018

References

- Ji, N. The practical and fundamental limits of optical imaging in mammalian brains. *Neuron* **83**, 1242–1245 (2014).
- Andersen, P., Morris, R., Amaral, D., Bliss, T. & O'Keefe, J. *The Hippocampus Book*. (Oxford University Press, Oxford, New York, 2007).
- Jones, E. G. *The Thalamus*. (Cambridge University Press, Cambridge, 2007).
- Misgeld, T. & Kerschensteiner, M. In vivo imaging of the diseased nervous system. *Nat. Rev. Neurosci.* **7**, 449–463 (2006).
- Čižmár, T. & Dholakia, K. Shaping the light transmission through a multimode optical fibre: complex transformation analysis and applications in biophotonics. *Opt. Express* **19**, 18871–18884 (2011).
- Čižmár, T. & Dholakia, K. Exploiting multimode waveguides for pure fibre-based imaging. *Nat. Commun.* **3**, 1027 (2012).
- Plöschner, M. & Čižmár, T. Compact multimode fiber beam-shaping system based on GPU accelerated digital holography. *Opt. Lett.* **40**, 197–200 (2015).
- Dombeck, D. A., Harvey, C. D., Tian, L., Looger, L. L. & Tank, D. W. Functional imaging of hippocampal place cells at cellular resolution during virtual navigation. *Nat. Neurosci.* **13**, 1433–1440 (2010).
- Szabo, V., Ventalon, C., De Sars, V., Bradley, J. & Emiliani, V. Spatially selective holographic Photoactivation and functional fluorescence imaging in freely behaving mice with a fiberscope. *Neuron* **84**, 1157–1169 (2014).
- Barretto, R. P. J. et al. Time-lapse imaging of disease progression in deep brain areas using fluorescence microendoscopy. *Nat. Med.* **17**, 223–228 (2011).
- Resendez, S. L. et al. Visualization of cortical, subcortical and deep brain neural circuit dynamics during naturalistic mammalian behavior with head-mounted microscopes and chronically implanted lenses. *Nat. Protoc.* **11**, 566–597 (2016).
- Bocarsly, M. E. et al. Minimally invasive microendoscopy system for in vivo functional imaging of deep nuclei in the mouse brain. *Biomed. Opt. Express* **6**, 4546–4556 (2015).
- Xu, H. T., Pan, F., Yang, G. & Gan, W. B. Choice of cranial window type for in vivo imaging affects dendritic spine turnover in the cortex. *Nat. Neurosci.* **10**, 549–551 (2007).
- Moshayedi, P. et al. The relationship between glial cell mechanosensitivity and foreign body reactions in the central nervous system. *Biomaterials* **35**, 3919–3925 (2014).
- Popoff, S. M. et al. Measuring the transmission matrix in optics: an approach to the study and control of light propagation in disordered media. *Phys. Rev. Lett.* **104**, 100601 (2010).
- Mahalati, R. N., Gu, R. Y. & Kahn, J. M. Resolution limits for imaging through multi-mode fiber. *Opt. Express* **21**, 1656–1668 (2013).
- Kim, G. et al. Deep-brain imaging via epi-fluorescence computational cannula microscopy. *Sci. Rep.* **7**, 44791 (2017).
- Feng, G. et al. Imaging neuronal subsets in transgenic mice expressing multiple spectral variants of GFP. *Neuron* **28**, 41–51 (2000).
- Crowe, S. E. & Ellis-Davies, G. C. R. Longitudinal in vivo two-photon fluorescence imaging. *J. Comp. Neurol.* **522**, 1708–1727 (2014).
- Ohayon, S., Caravaca-Aguirre, A., Piestun, R. & DiCarlo, J. J. Minimally invasive multimode optical fiber microendoscope for deep brain fluorescence imaging. *Biomed. Opt. Express* **9**, 1492–1509 (2018).
- Plöschner, M., Straka, B., Dholakia, K. & Čižmár, T. GPU accelerated toolbox for real-time beam-shaping in multimode fibres. *Opt. Express* **22**, 2933–2947 (2014).
- Conkey, D. B., Caravaca-Aguirre, A. M. & Piestun, R. High-speed scattering medium characterization with application to focusing light through turbid media. *Opt. Express* **20**, 1733–1740 (2012).
- Mitchell, K. J., Turtaev, S., Padgett, M. J., Čižmár, T. & Phillips, D. B. High-speed spatial control of the intensity, phase and polarisation of vector beams using a digital micro-mirror device. *Opt. Express* **24**, 29269–29282 (2016).
- Turtaev, S. et al. Comparison of nematic liquid-crystal and DMD based spatial light modulation in complex photonics. *Opt. Express* **25**, 29874–29884 (2017).
- Gigan, S. Optical microscopy aims deep. *Nat. Photon.* **11**, 14–16 (2017).
- Loterie, D. et al. Digital confocal microscopy through a multimode fiber. *Opt. Express* **23**, 23845–23858 (2015).
- Morales-Delgado, E. E., Farahi, S., Papadopoulos, I. N., Psaltis, D. & Moser, C. Delivery of focused short pulses through a multimode fiber. *Opt. Express* **23**, 9109–9120 (2015).
- Sahl, S. J., Hell, S. W. & Jakobs, S. Fluorescence nanoscopy in cell biology. *Nat. Rev. Mol. Cell Biol.* **18**, 685–701 (2017).